

Bone&JointEvidence

Your Topics in Focus

Vol 01, No 02
October 2011

How to Achieve Alignment in Total Knee Arthroplasty

A Systematic Review and Meta-Analysis

Reviewed by: **Henrik Schroeder-Boersch** and **Friedrich Boettner**.



Use this QR code to access the main Bone&JointEvidence review online.

Authors:

Shelly-Ann Rampersad BSc¹, Nikhil Kitchlu BSc¹, Mike Saccone BSc¹,
Sheila Sprague MSc^{1,2}, Mohit Bhandari MD, PhD, FRCSC^{1,2}

- 1 Global Research Solutions™
- 2 Centre for Evidence-Based Orthopaedics

Reviewers:

Henrik Schroeder-Boersch MD¹, Friedrich Boettner MD²

- 1 Aukamm-Klinik, Wiesbaden, Germany
- 2 Hospital for Special Surgery, New York, NY, USA

Table of contents

OVERVIEW	Overview	3
BACKGROUND	Background	4
METHODS	Methods	5
RESULTS	Results	6
CONCLUSIONS	Conclusions	15
REVIEW	Review	16
REFERENCES	References	17
APPENDICES	Appendices	17


Overview

How to Achieve Alignment in Total Knee Arthroplasty: A Systematic Review and Meta-Analysis.

Purpose of review The purpose of the current study was to conduct a thorough systematic literature review and meta-analysis of tools and methods used to achieve alignment in total knee arthroplasty (TKA).

Background An important consideration with total knee arthroplasty is whether the neutral alignment of the leg will be correctly restored following surgery. We performed a systematic literature review of different methods of total knee arthroplasty to improve the understanding of alignment outcomes to date. [Read more on page 4](#)

Rating



Why this rating? This is a systematic review and meta-analysis of exclusively current Level I evidence. The evidence rating is very good.

Results Twenty-nine studies which included data from 3,354 total knee arthroplasties met the inclusion criteria for this review. Data were pooled based on four different groups comparing:

- (1) imageless navigation and conventional TKA;
- (2) image-based navigation and conventional TKA;
- (3) intramedullary and extramedullary alignment;
- (4) different navigation systems.

[Read more on pages 6–14](#)

Key considerations

A systematic review of the literature found:

- **There was a relative risk reduction of outliers** (alignment greater than +/- 3 degrees from neutral) **in imageless navigation compared to conventional TKA** (63%, 55% and 57% in the mechanical axis, femoral angle and tibial angle alignment, respectively).
- **Meta-analytic techniques could not be used** **in the remaining comparison groups due to limited data available.**
- **Need for additional studies with:**
 - Larger sample sizes
 - Consistent reporting amongst studies
 - Comparisons between image-based navigation and conventional systems
 - Comparisons between different types of navigation

Background

Total knee arthroplasty (TKA) is a well accepted treatment for osteoarthritis of the knee [1]. An important consideration with TKA is whether the neutral alignment of the leg will be correctly restored following surgery. Malalignment can lead to increased risk of loosening, instability, pain and subsequently, the need for revision [2]. Various systems exist that are utilized for the restoration of alignment, including conventional systems such as intramedullary and extramedullary guides and cutting blocks, as well as more modern computer navigation systems with and without imaging. Whether or not these truly improve the accuracy of alignment is still a matter of debate within the literature [3].


We performed a systematic review and meta-analysis of exclusively randomized controlled trials (level I evidence) concerning alignment in total knee arthroplasty. Studies using imageless computer navigation, image-based computer navigation, and intramedullary and extramedullary alignment for TKA were included. Primary outcomes were the percentage of outliers, defined as alignment deviations greater than three degrees from neutral, in the mechanical axis, femoral angle, and tibial angle. These outcomes were evaluated in four main groups of patients in which the methods compared were: (1) imageless computer navigation versus conventional TKA; (2) image-based navigation versus conventional TKA; (3) intramedullary versus extramedullary alignment; and (4) different types of navigation.

Methods

From 646 potentially eligible studies identified by a literature search, 617 did not meet the eligibility criteria for this review, leaving 29 eligible studies (**Figure 1**).

Please refer to *Appendix 1: Methods* for further detail on the eligibility criteria and literature search.

Figure 1: Literature Review



Study Characteristics

Study characteristics are summarized in **Figure 2** with further detail found in **Tables 1 to 4**.

Please refer to *Appendix 2: Results* for further detail on the study characteristics.

Figure 2: Study characteristics

		Comparison of imageless navigation to conventional TKA		Comparison of image-based navigation to conventional TKA	
		Navigation	Conventional	Navigation	Conventional
	Study designs included:	Randomized studies	Randomized studies	Randomized studies	Randomized studies
	Mean age:	68.3 years	68.6 years	72.0 years	70.0 years
	Mean sample size:	62.4	57.9	50.0	50.0
	% male:	25.5	25.9	35.1	35.1
	Number of knees in study:	3008 (25 studies)		100 (1 study)	
		Comparison of intramedullary to extramedullary alignment		Comparison of two different types of navigation	
		Intramedullary	Extramedullary	Navigation	Conventional
	Study designs included:	Randomized studies	Randomized studies	Randomized studies	Randomized studies
	Mean age:	70.0 years	69.0 years	71.3 years	65.2 years
	Mean sample size:	52.0	48.0	27.0	19.0
	% male:	42.5	44.1	33.0	47.0
	Number of knees in study:	200 (2 studies)		46 (1 study)	

Results Cont.

The percentage of outliers in the mechanical axis, femoral angle and tibial angle for leg alignment are summarized in **Tables 1 to 4**.

Findings are indicative of a reduced rate of alignment outlier when imageless navigation is used in comparison to conventional systems across all three outcomes.

The pooled relative risk of an outlier for mechanical axis, tibial angle, and femoral angle were as follows (**Tables 5 to 7**):

- Mechanical axis: 0.37 (95% CI: 0.28 to 0.50; p <0.001)
- Tibial angle: 0.45 (95% CI: 0.33 to 0.61; p <0.001)
- Femoral angle: 0.43 (95%CI: 0.31 to 0.59; p <0.001)

Unfortunately, due to insufficient data pooled relative risks for the other comparison groups could not be determined.

Please refer to *Appendix 2*: Results for additional details on the study results.

Table 1: Characteristics of included studies comparing imageless navigation to conventional TKA

Study	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Mean	Navigation Conventional	62.4 57.9	68.3 68.6	25.5 25.9	8.9 25.9	6.4 15.5	5.0 13.5	
Bathis et al, 2004	BrainLab navigation	80	68.7	26.3	4.0	8.0	2.0	NR
	Conventional PFC Sigma	80	70.9	33.8	22.0	14.0	6.0	
Bohling et al, 2005	Surgetics navigation	50	69.0	28.0	6.0	NR	NR	NR
	Technical instrumentation	50	70.0	22.0	54.0	NR	NR	
Chauhan et al, 2004	Stryker navigation	35	NR	NR	14.3	0.0	0.0	NR
	Conventional jig-based technique	35	NR	NR	28.6	8.6	8.6	
Choong et al, 2009	Ci Navigation, Depuy	60	70.0	29.8	12.0	NR	NR	Osteoarthritis (103) Rheumatoid arthritis (8)
	Conventional approach using intramedullary and extramedullary guides (Sigma SP2)	55	69.0	50.0	39.0	NR	NR	
Chotanaphuti et al, 2008	BrainLab navigation	86	67.8	11.7	6.0	7.0	3.5	Osteoarthritis (168) Rheumatoid arthritis (12)
	Conventional approach using intramedullary and extramedullary guides	94	67.2	7.0	13.0	14.9	8.5	
Confalonieri et al, 2007	BrainLab navigation	37	73.4	49.0	0.0	5.4	0.0	NR
	Minimally invasive approach using intramedullary and extramedullary guides	37	72.8	46.0	16.2	13.5	16.2	

NR Not reported

Table 1: Characteristics of included studies comparing imageless navigation to conventional TKA (Cont.)

Study	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Decking et al, 2005	OrthoPilot navigation	27	64.7	33.0	48.1§	18.5§	3.7§	Primary osteoarthritis (39) Rheumatoid arthritis (5) Posttraumatic osteoarthritis (8)
	Manual	25	67.3	32.0	64§	20.0§	20.0§	
Ensini et al, 2006	Stryker navigation	60	68.8	50.0	11.7	0.0	1.7	Primary arthritis secondary to articular or extraarticular fractures and avascular necrosis (120)
	Conventional	60	71.1	33.0	25.0	15.0	3.3	
Hernandez-Vaquero et al, 2010	Stryker navigation	20 OA 20 PD	NR	NR	0.0 OA 10.0 PD	NR	NR	Osteoarthritis (40) Preoperative deformity
	Conventional approach using intramedullary and extramedullary guides	20 OA 20 PD	NR	NR	22.5 OA 50.0 PD	NR	NR	
Hsu et al, 2010	BrainLab navigation	60	70.0	32.0	13.3	10.0	1.7	Bilateral osteoarthritis (NR)
	Conventional approach using intramedullary and extramedullary guides	60			50.0	13.3	11.7	
Kim et al, 2007	BrainLab navigation	100	67.6	15.0	21.0	NR	NR	Bilateral Osteoarthritis (200)
	Conventional approach using intramedullary and extramedullary guides	100			18.0	NR	NR	
Lee et al, 2010	OrthoPilot navigation	60	66.0	5.0	8.0	NR	NR	Osteoarthritis (116)
	Conventional measured resection technique	56	67.0	3.6	27.0	NR	NR	
Luring et al, 2008	BrainLab navigation and minimal invasive surgical (MIS) technique with special MIS pathway instruments	30	70.0	NR	0.0	0.0	0.0	NR
	Conventional technique using the standard PFC specialist II instruments.	30	69.0	NR	5.0	0.0	3.3	
	Freehand MIS with special MIS pathway instruments	30	69.0	NR	10.0	3.3	3.3	
Lutzner et al, 2008	Stryker navigation	40	69.0	32.5	13.0	15.0	NR	Primary or secondary osteoarthritis (80)
	Conventional approach using intramedullary and extramedullary guides	40	69.0	40.0	18.0	5.0	NR	

NR Not reported
OA Osteoarthritis

PD Post-operative deformity
§ Percentages reported include 3°

Table 1: Characteristics of included studies comparing imageless navigation to conventional TKA (Cont.)

Study	Randomized studies Prospective studies Case series	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Martin et al, 2007		CT-free VectorVision from BrainLAB	100	70.3	32.0	8.0	5.0	2.0	Osteoarthritis (198) Rheumatoid arthritis (1) Ahlback's disease (1)
		Conventional approach using intramedullary and extramedullary guides	100	71.1	27.0	24.0	14.0	20.0	
Matsumoto et al, 2004		CT-free VectorVision from Brainlab	30	75.3	16.7	12.5	12.5	12.5	Osteoarthritis (60)
		Conventional	30	73.3	16.7	37.5	34.4	28.1	
Matziolis et al, 2007		PIGalileo navigation	32	71.0	33.3	3.1	0.0	0.0	Primary arthritis of the knee (60)
		Conventional technique using intramedullary guide	28	70.0		25.0	10.7	17.9	
Mombert et al, 2007		VectorVision TKR navigation system from Brainlab	21	NR	NR	0.0	NR	NR	NR
		Conventional technique using intramedullary guiding system	21	NR	NR	19.0	NR	NR	
Mullaji et al, 2007		Ci navigation, BrainLab	282	65.5	24.0	9.2	9.21*	9.21*	NR
		Image intensifier-guided conventional TKA	185	65.9	23.0	21.6	23.8*	23.8*	
Oberst et al, 2007		VectorVision system from BrainLAB	34	NR	NR	6.0	NR	NR	NR
		Conventional approach using intramedullary and extramedullary guides	35	NR	NR	20.0	NR	NR	
Schmitt et al, 2011		NexGen LPS with navigation	30	70.2	26.7	NR	0.0	12.3	Gonarthrosis (90)
		Stryker navigation- Scorpio PS	30	69.2	26.7	NR	0.0	6.7	
		Conventional NexGen LPS	30	69.6	40.0	NR	10.0	3.3	
Seon et al, 2006		OrthoPilot (Aesculap) – Navigation-assisted less invasive surgical technique (LIS)	49	67.4	10.2	6.1	6.1	4.1	Osteoarthritis (101) Rheumatoid arthritis (1)
		Conventional approach using intramedullary and extramedullary guides	53	64.8	11.3	11.3	17.0	7.5	

NR Not reported

* Outliers for femoral and tibial angles reported together

Table 1: Characteristics of included studies comparing imageless navigation to conventional TKA (Cont.)

Study	Randomized studies Prospective studies Case series	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Sparmann et al, 2003		Stryker navigation	120	67.4	26.7	0.0	0.8	0.8	Primary osteoarthritis (142) Post-traumatic osteoarthritis (7) Rheumatoid arthritis (65) Psoriatic arthropathy (11) Spondyloarthropathy (8) Scleroderma (2) Lupus erythematosus (5)
		Conventional hand-guided technique	120	66.1	34.2	13.3	28.3	10.0	
Stockl et al, 2004		Stryker navigation	32	68.2	28.1	0.0	NR	NR	Osteoarthritis (64)
		Conventional method using extramedullary alignment guides	32	72.4	31.3	6.3	NR	NR	
van Strien et al, 2009		BrainLab navigation – CT-based	17	71.0	NR	29.4	23.5	23.5	NR
		Brainlab navigation – CT-free	19		NR	10.5	0.0	26.3	
		Conventional	21		NR	33.3	4.8	42.9	

NR Not reported

Table 2: Characteristics of included studies comparing image-based navigation to conventional TKA

Study	Randomized studies Prospective studies Case series	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Mean		Navigation	50.0	72.0	35.1	0.0	NR	NR	
		Conventional	50.0	70.0	35.1	26.5	NR	NR	
Victor et al, 2004		Fluoroscopy-based navigation	50	72.0	35.1	0.0	NR	NR	NR
		Conventional approach using intramedullary and extramedullary guides	50	70.0	35.1	26.5	NR	NR	

NR Not reported

Table 3: Characteristics of included studies comparing intramedullary to extramedullary alignment

Study	Randomized studies Prospective studies Case series	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Mean			52.0 48.0	70.0 69.0	42.5 44.1	NR NR	10.0 6.0	15.0 35.0	
Baldini et al, 2008		Intramedullary device	50	71.0	33.0	NR	10.0	NR	Varus osteoarthritis (100)
		Extramedullary device	50	70.0	37.0	NR	6.0	NR	
Reed et al, 2002		Intramedullary device	54	69.0	51.9	NR	NR	15.0	Osteoarthritis (93) Inflammatory arthritis (7)
		Extramedullary device	46	68.0	52.2	NR	NR	35.0	

NR Not reported

Table 4: Characteristics of included studies comparing two different types of navigation

Study	Randomized studies Prospective studies Case series	System Used	Sample Size (Number of Knees)	Mean Age (Years)	% Male	Mechanical axis > ±3° (%)	Femoral angle > ±3° (%)	Tibial angle > ±3° (%)	Indication for Procedure
Mean			27.0 19.0	71.3 65.2	33.0 47.0	0.0 5.3	NR NR	NR NR	
Lionberger et al, 2008		Electromagnetic navigation system	27	71.3	33.0	0.0	NR	NR	Osteoarthritis (NR) Rheumatoid arthritis (NR)
		Infrared navigation system	19	65.2	47.0	5.3	NR	NR	

NR Not reported

Table 5: Forest plot of relative risk for the outliers in the mechanical axis for imageless navigation vs. conventional.

Study or Subgroup	Navigation		Conventional		Weight	Risk Ratio M-H, Random, 95% CI	Risk Ratio M-H, Random, 95% CI
	Events	Total	Events	Total			
Bathis 2004	3	80	18	80	3.8%	0.17 [0.05, 0.54]	
Bohling 2005	3	50	27	50	4.0%	0.11 [0.04, 0.34]	
Chauhan 2004	5	35	10	35	4.8%	0.50 [0.19, 1.31]	
Choong 2009	7	60	21	55	6.0%	0.31 [0.14, 0.66]	
Chotanaphuti 2008	5	86	12	94	4.6%	0.46 [0.17, 1.24]	
Confalonieri 2007	0	37	6	37	0.9%	0.08 [0.00, 1.32]	
Decking 2005	13	27	16	25	8.1%	0.75 [0.46, 1.23]	
Ensini 2007	7	60	15	60	5.7%	0.47 [0.20, 1.06]	
Hernandez-Vaquero 2010	2	40	15	40	3.0%	0.13 [0.03, 0.55]	
Hsu 2010	8	60	30	60	6.6%	0.27 [0.13, 0.53]	
Kim 2007	21	100	18	100	7.5%	1.17 [0.66, 2.05]	
Lee 2010	5	60	16	56	5.0%	0.29 [0.11, 0.74]	
Luring 2008	0	30	5	60	0.9%	0.18 [0.01, 3.13]	
Lutzner 2010	5	40	7	40	4.3%	0.71 [0.25, 2.06]	
Martin 2007	8	100	24	100	6.2%	0.33 [0.16, 0.71]	
Matsumoto 2004	4	30	11	30	4.5%	0.36 [0.13, 1.01]	
Matziolis 2007	1	32	7	28	1.7%	0.13 [0.02, 0.95]	
Mombert 2007	0	21	4	21	0.9%	0.11 [0.01, 1.94]	
Mullaji 2007	26	282	40	185	8.3%	0.43 [0.27, 0.67]	
Oberst 2008	2	34	7	35	2.7%	0.29 [0.07, 1.32]	
Seon 2006	3	49	6	53	3.3%	0.54 [0.14, 2.05]	
Sparmann 2003	0	120	16	120	1.0%	0.03 [0.00, 0.50]	
Stockl 2004	0	32	2	32	0.9%	0.20 [0.01, 4.01]	
van Strien 2009	7	36	7	21	5.2%	0.58 [0.24, 1.43]	
Total (95% CI)		1501	1417		100.0%	0.37 [0.28, 0.50]	
Total Events	135		340				


Heterogeneity: $Tau^2 = 0.21$; $Chi^2 = 44.19$, $df = 23$ ($P = 0.005$); $I^2 = 48\%$
 Test for overall effect: $Z = 6.69$ ($P < 0.00001$)

0.001 0.1 1 10 1000
 Favors navigation Favors conventional

CI Confidence interval
 M-H, Random A Mantel-Haenszel, random-effects model was used to pool the data.

◆ Diamonds indicate pooled relative risk estimates.
 ■ Squares indicate point estimates around which 95% confidence intervals are denoted by a horizontal line.

Table 6: Forest plot of relative risk for the outliers in femoral angle for imageless navigation vs. conventional.




Heterogeneity: $Tau^2 = 0.05$; $Chi^2 = 18.48$, $df = 16$ ($P = 0.030$); $I^2 = 13\%$
 Test for overall effect: $Z = 5.29$ ($P < 0.00001$)

0.001 0.1 1 10 1000
 Favors navigation Favors conventional

CI Confidence interval
 M-H, Random A Mantel-Haenszel, random-effects model was used to pool the data.

◆ Diamonds indicate pooled relative risk estimates.
 ■ Squares indicate point estimates around which 95% confidence intervals are denoted by a horizontal line.

Table 7: Forest plot of relative risk for the outliers in the tibial angle for imageless navigation vs. conventional.



CI Confidence interval
 M-H, Random A Mantel-Haenszel, random-effects model was used to pool the data.

◆ Diamonds indicate pooled relative risk estimates.
 ■ Squares indicate point estimates around which 95% confidence intervals are denoted by a horizontal line.

Conclusions

This systematic review aims to assess the degree of realignment in TKA patients, comparing:

- 1) imageless navigation to conventional TKA,
- 2) image-based navigation to conventional TKA,
- 3) intramedullary and extramedullary alignment, and
- 4) different types of navigation systems.

Specifically, the post-operative leg alignments were analyzed to determine the percentage of outliers outside the recommended $\pm 3^\circ$ varus/valgus range. Evidence from our meta-analysis is indicative of a significant reduction in the relative risk of an outlier alignment when imageless computer navigation systems are used in comparison to conventional methods. Conventional methods included intramedullary and extramedullary guides. The lack of high quality data available prevented any conclusions from being established in the other aforementioned comparisons. Additional randomized controlled trials comparing image-based navigation and conventional TKA, and between different types of navigation is required to improve the quality of the literature available.

Strengths

- A thorough and systematic review of the literature was conducted using multiple databases.
- Demonstrated reproducibility of selection and quality of assessment criteria.
- The inclusion of only high quality evidence (Level I randomized controlled trials), allowing for meta-analysis.
- High degree of external validity with data from over 3,000 male and female knees that underwent total knee arthroplasty in 29 different studies from a variety of countries.
- Meta-analytic techniques combined high quality evidence from 25 randomized controlled trials comparing imageless computer navigation and conventional methods for achieving knee alignment after TKA, indicating a clear benefit when navigation is used.

Limitations

- Limited studies comparing image-based navigation with conventional TKA, intramedullary and extramedullary alignment and comparison of different navigation systems.
- Limited reporting of specific outcomes in studies.
- Heterogeneity as a result of the different systems and measurement techniques used in the studies, decreased the validity of findings.
- Inability to determine clinical importance of the relative risk reduction in outliers provided by imageless computer navigation in comparison to conventional methods.

Review at a glance

Generalizability

90 out of 100. The included studies assess methods for total knee arthroplasty for any indication, allowing the findings to be applied to a larger population with similar characteristics.

Validity

90 out of 100. Systematic review of high evidence, with fairly consistent reporting on outcomes between included studies.

Timeliness


90 out of 100. This review assesses several alternatives in total knee arthroplasty treatment used for varying indications. All studies in this review were published within the past nine years.

Importance

85 out of 100. The evidence is important in providing patients, orthopaedic surgeons, and healthcare payers information regarding the success in achieving desired alignment when using different systems and methods for total knee arthroplasty.

Strength

90 out of 100. Data from 29 randomized controlled trials were included in this study. The quality of evidence is high, and meta-analytic statistics were performed, allowing for reliable conclusions to be drawn.



References

1. Dieppe P, Basler H-D, Chard J et al. (1991) Knee replacement surgery for osteoarthritis: effectiveness, practice variations, indications and possible determinants of utilization. *Br Soc Rheumatology*;38:73-83
2. Jeffery RS, Morris RW, Denham RA. (1991 Sep) Coronal alignment after total knee replacement. *J Bone Joint Surg Br*.;73(5):709-14.
3. Matziolis G, Krockner D, Weiss U, Tohtz S, Perka C. (2007 Feb) A prospective, randomized study of computer-assisted and conventional total knee arthroplasty. Three-dimensional evaluation of implant alignment and rotation. *J Bone Joint Surg Am*.;89(2):236-43.
4. Bathis H, Perlick L, Tingart M et al. (2004 Feb) Alignment in total knee arthroplasty. A comparison of computer-assisted surgery with the conventional technique. *J Bone Joint Surg Br*.;86(3):372-7.
5. Bohling U, Schamberger H, Grittner U et al. (2005 Jun) Computerised and technical navigation in total knee-arthroplasty. *Journal of Orthopaedics and Traumatology*.;6(2):69-75.
6. Chauhan SK, Scott RG, Breidahl W et al. (2004 Apr) Computer-assisted knee arthroplasty versus a conventional jig-based technique. A randomised, prospective trial. *J Bone Joint Surg Br*.;86(3):372-7.
7. Choong PF, Dowsey MM, Stoney JD. (2009 Jun) Does accurate anatomical alignment result in better function and quality of life? comparing conventional and computer-assisted total knee arthroplasty. *J Arthroplasty*.;24(4):560-9.
8. Chotanaphuti T, Ongnamthip P, Teeraleekul K et al. (2008 Sep) Comparative study between computer assisted-navigation and conventional technique in minimally invasive surgery total knee arthroplasty, prospective control study. *J Med Assoc Thai*.;91(9):1382-8.
9. Confalonieri N, Manzotti A, Pullen C, Ragone V. (2007 Dec) Mini-incision versus mini-incision and computer-assisted surgery in total knee replacement: A radiological prospective randomised study. *Knee*.;14(6):443-7.
10. Decking R, Markmann Y, Fuchs J et al. (2005 Apr) Leg axis after computer-navigated total knee arthroplasty: A prospective randomized trial comparing computer-navigated and manual implantation. *J Arthroplasty*.;20(3):282-8.
11. Ensini A, Catani F, Leardini A et al. (2007 Apr) Alignments and clinical results in conventional and navigated total knee arthroplasty. *Clin Orthop*.;457:156-62.
12. Hernandez-Vaquero D, Suarez-Vazquez A, Sandoval-Garcia MA et al. (2010 May) Computer assistance increases precision of component placement in total knee arthroplasty with articular deformity. *Clin Orthop*.;468(5):1237-41.
13. Hsu WH, Hsu RW, Weng YJ. (2010 Oct) Effect of preoperative deformity on postoperative leg axis in total knee arthroplasty: A prospective randomized study. *Knee Surg Sports Traumatol Arthrosc*.;18(10):1323-7.
14. Kim YH, Kim JS, Yoon SH. (2007 Apr) Alignment and orientation of the components in total knee replacement with and without navigation support: A prospective, randomised study. *J Bone Joint Surg Br*.;89(4):471-6.
15. Lee DH, Park JH, Song DI et al. (2010 Mar) Accuracy of soft tissue balancing in TKA: Comparison between navigation-assisted gap balancing and conventional measured resection. *Knee Surg Sports Traumatol Arthrosc*.;18(3):381-7.
16. Luring C, Beckmann J, Haibock P, et al. Minimal invasive and computer assisted total knee replacement compared with the conventional technique: A prospective, randomised trial. *Knee Surg Sports Traumatol Arthrosc*. 2008 Oct;16(10):928-34.
17. Lutzner J, Gunther KP, Kirschner S. (2010 Oct) Functional outcome after computer-assisted versus conventional total knee arthroplasty: A randomized controlled study. *Knee Surg Sports Traumatol Arthrosc*.;18(10):1339-44.
18. Martin A, Wohlgenannt O, Prens M et al. (2007 Jul) Imageless navigation for TKA increases implantation accuracy. *Clin Orthop*. 2007(460):178-84.
19. Matsumoto T, Tsumura N, Kurosaka M et al. (2004) Prosthetic alignment and sizing in computer-assisted total knee arthroplasty. *Int Orthop*.;28:282-5.
20. Mombert M, Van Den Daelen L, Gunst P et al. (2007 Feb) Navigated total knee arthroplasty: A radiological analysis of 42 randomised cases. *Acta Orthop Belg*.;73(1):49-54.
21. Mullaji A, Kanna R, Marawar S et al. (2007 Oct) Comparison of limb and component alignment using computer-assisted navigation versus image intensifier-guided conventional total knee arthroplasty: A prospective, randomized, single-surgeon study of 467 knees. *J Arthroplasty*.;22(7):953-9.
22. Oberst M, Bertsch C, Konrad G et al. (2008 Jun) CT analysis after navigated versus conventional implantation of TKA. *Arch Orthop Trauma Surg*.;128(6):561-6.
23. Schmitt J, Hauk C, Kienapfel H et al. (2011) Navigation of total knee arthroplasty: Rotation of components and clinical results in a prospectively randomized study. *BMC Musculoskeletal Disorders*. 2011;12.
24. Seon JK, Song EK. (2006 Sep) Navigation-assisted less invasive total knee arthroplasty compared with conventional total knee arthroplasty: A randomized prospective trial. *J Arthroplasty*.;21(6):777-82.
25. Sparmann M, Wolke B, Czupalla H et al. (2003 Aug) Positioning of total knee arthroplasty with and without navigation support. A prospective, randomised study. *J Bone Joint Surg Br*.;85(6):830-5.
26. Stockl B, Nogler M, Rosiek R et al. (2004 Sep) Navigation improves accuracy of rotational alignment in total knee arthroplasty. *Clin Orthop*. (426):180-6.
27. van Strien T, van der Linden-van der Zwaag E., Kaptein B et al. (2009 Oct) Computer assisted versus conventional cemented total knee prostheses alignment accuracy and micromotion of the tibial component. *Int Orthop*.;33(5):1255-61.
28. Victor J, Hoste D. (2004 Nov) Image-based computer-assisted total knee arthroplasty leads to lower variability in coronal alignment. *Clin Orthop*. (428):131-9.
29. Baldini A, Adravanti P. (2008 Nov) Less invasive TKA: Extramedullary femoral reference without navigation. *Clin Orthop*.;466(11):2694-700.
30. Reed MR, Bliss W, Sher JL et al. (2002 Aug) Extramedullary or intramedullary tibial alignment guides: A randomised, prospective trial of radiological alignment. *J Bone Joint Surg Br*.;84(6):858-60.
31. Lionberger DR, Weise J, Ho DM et al. (2008 Jun) How does electromagnetic navigation stack up against infrared navigation in minimally invasive total knee arthroplasties?. *J Arthroplasty*.;23(4):573-80.



Appendices

Visit www.kleos.md/literature/bone-joint-evidence/appendices/v01/n02.pdf for appendices or use this QR code. The following appendices provide further detail:

Appendix 1: Methods

Appendix 2: Results

Please also visit www.kleos.md/literature/background to view an online glossary of technical terminology.

Notes

A series of horizontal dotted lines for taking notes.

Disclaimer Great care has been taken to maintain the accuracy of the information contained in the publication. However, neither KLEOS, nor the authors can be held responsible for errors or any consequences arising from the use of the information contained in this publication. The statements or opinions contained in editorials and articles in this journal are solely those of the authors thereof and not of KLEOS. The products, procedures, and therapies described are only to be applied by certified and trained medical professionals in environments specially designed for such procedures. No suggested test or procedure should be carried out unless, in the reader's professional judgment, its risk is justified. Because of rapid advances in the medical sciences, we recommend that independent verification of diagnosis, drugs dosages, and operating methods should be made before any action is taken. Although all advertising material is expected to conform to ethical (medical) standards, inclusion in this publication does not constitute a guarantee or endorsement of the quality or value of such product or of the claims made of it by its manufacturer. Some of the products, names, instruments, treatments, logos, designs, etc. referred to in this journal are also protected by patents and trademarks or by other intellectual property protection laws even though specific reference to this fact is not always made in the text. Therefore, the appearance of a name, instrument, etc. without designation as proprietary is not to be construed as a representation by the publisher that it is in the public domain. This publication, including all parts thereof, is legally protected by copyright. Any use, exploitation or commercialization outside the narrow limits of copyrights legislation, without the publisher's consent, is illegal and liable to prosecution. This applies in particular to photostat reproduction, copying, scanning or duplication of any kind, translating, preparation of microfilms and electronic data processing and storage. Institutions' subscriptions allow to reproduce tables of content or prepare lists of articles including abstracts for internal circulation within the institutions concerned. Permission of the publisher is required for resale or distribution outside the institutions. Permission of the publisher is required for all other derivative works, including compilations and translations. Permission of the publisher is required to store or use electronically any material contained in this journal, including any article or part of an article. For inquiries contact the publisher at the address indicated.

US Lit No: 71281876 REVO 9/11

Produced and published by KLEOS, the medical education service from Smith & Nephew.

Published October 2011

Copyright © 2011 by Smith & Nephew, KLEOS medical education.

All rights reserved. KLEOS is a trademark of Smith & Nephew.

KLEOS, Oberneuhofstrasse 10d, 6340 Baar, Switzerland

Phone +41 41 766 22 55

kleos@smith-nephew.com

Bone&JointEvidence

is available on the KLEOS website within "Literature"

Come and visit us at www.kleos.md